

Optimal Sensor Design Using Patient-Specific Images

Sukhi Basati, Timothy J. Harris, Jr., and Andreas A. Linninger
University of Illinois at Chicago

In diseases such as hydrocephalus, the cerebral ventricles enlarge. The treatment options for these patients are presently based on pressure, which has limited capabilities. We present the design of a volume sensor as an alternative monitoring option. Through the use of computer aided design and simulation, we optimized a sensor in silico with fewer resources. Specifically, we designed a

sensor for animal experimentation with a scalable procedure for human sensors. In this paper, we present a rational design approach for a sensor that integrates advances in medical imaging. Magnetic resonance data sets of both normal and diseased subjects were used as a virtual laboratory. Finite element simulations were performed under pathological disease states of the brain as a contribution toward an accelerated device design. An optimized sensor was then fabricated for these subjects based on the outcome of the simulations. In this paper, we explain how a computer aided subject-specific design was used to help fabricate and test our sensor.

Statistical Shape Modeling of Femurs Using Morphing and Principal Component Analysis

Najah Hraiech, Christelle Boichon, Michel Rochette, Thierry Marchal, and Marc Horner
ANSYS, Inc.

In this paper, we describe a method for automatically building a statistical shape model by applying a morphing method and a principal component analysis (PCA) to a large database of femurs. One of the major challenges in building a shape model from a training data set of 3D objects is the determination of the correspondence between different shapes. In our work, we solve this

problem by using a morphing method. The morphing method consists of deforming the same template mesh over a large database of femur geometries, which results in isotopological meshes and one to one correspondences; i.e., the resulting meshes have the same number of nodes, the same number of elements, and the same connectivity in all morphed meshes. By applying the morphing-based registration followed by PCA to a large database of femurs, we demonstrate that the method can be used to derive a low dimensional representation of the main variabilities of the femur geometry.